

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q2125 SAS No.: Q2125 SDG No.: Q2125

Instrument ID: BNA_P Calibration Date/Time: 06/11/2025 10:09

Lab File ID: BP024905.D Init. Calib. Date(s): 06/06/2025 06/06/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:30 15:18

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.198	1.226		2.3	
Phenol-d6	1.585	1.563		-1.4	
Nitrobenzene-d5	0.412	0.408		-1.0	
Naphthalene	1.025	1.019		-0.6	
2-Methylnaphthalene	0.650	0.650		0.0	
2-Fluorobiphenyl	1.485	1.486		0.1	
2,4,6-Tribromophenol	0.277	0.297		7.2	
Terphenyl-d14	1.116	1.127		1.0	

All other compounds must meet a minimum RRF of 0.010.