

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: ALLI03
 ProjectID: LOD-LOQ-MDL Study
 Lab Code: CHEM Case No.: Q2126 SAS No.: Q2126 SDG No.: Q2126

Calibration Sequence : FF052125			Test : Diesel Range Organics	
Concentration	(PPM)	Area Count	Reference Factor	File ID
1000		131501360	131501	FF015874.D
500		66823160	133646	FF015875.D
200		25639665	128198	FF015876.D
100		14440283	144403	FF015877.D
50		6491363	129827	FF015878.D
AVG RF : 133515			% RSD : 4.803	AVG RT : 15.0206