

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: ALLI03
 ProjectID: LOD-LOQ-MDL Study
 Lab Code: CHEM Case No.: Q2126 SAS No.: Q2126 SDG No.: Q2126

Calibration Sequence : FG052225		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
1000	132190666	132191	FG015886.D
500	67168968	134338	FG015887.D
200	26324104	131621	FG015888.D
100	13743634	137436	FG015889.D
50	6335792	126716	FG015890.D
AVG RF : 132460		% RSD : 2.974	AVG RT : 14.9848