

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark		Date Collected:	05/23/25	
Project:	LOD-LOQ-MDL Study		Date Received:	05/23/25	
Client Sample ID:	LOD-MDL-WATER-01-QT2-2025		SDG No.:	Q2126	
Lab Sample ID:	Q2126-07		Matrix:	Water	
Analytical Method:	8015D DRO		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1	mL
Soil Aliquot Vol:		uL	Test:	Diesel Range Organics	
Extraction Type:			Injection Volume :		
GPC Factor :		PH :			
Prep Method :	SW3510				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
FF015939.D	1	06/10/25 08:00	06/11/25 2:23	PB168381

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
DRO	DRO	15.0	J	6.00	50.0	ug/L
SURROGATES						
16416-32-3	Tetracosane-d50	17.4		29 - 130	87%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit