

Continuing Calibration Report for SequenceID : FC061125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069180.D

Aliphatic C9-C12	5744003.000	60.000	3.303	6.605	95733.383	105717.491	9.444
Aliphatic C12-C16	3539836.000	40.000	6.606	10.010	88495.900	101985.435	13.227
Aliphatic C16-C21	4981992.000	60.000	10.011	13.380	83033.200	97865.081	15.155
Aliphatic C21-C28	6974670.000	80.000	13.381	17.048	87183.375	93803.915	7.058
Aliphatic C28-C40	10669142.000	120.000	17.049	22.033	88909.517	94439.035	5.855
Aliphatic EPH	31909643.000	360.000	3.303	22.033	88637.897	97587.136	9.171