

Continuing Calibration Report for SequenceID : FD061125AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049455.D

Aromatic C10-C12	6607565.000	40.000	4.424	6.168	165189.125	167344.833	1.288
Aromatic C12-C16	10601544.000	60.000	6.169	8.797	176692.400	187525.524	5.777
Aromatic C16-C21	13766641.000	80.000	8.798	13.088	172083.013	184740.440	6.851
Aromatic C21-C36	27917048.000	180.000	13.089	18.509	155094.711	145057.816	-6.919
Aromatic EPH	58892798.000	360.000	4.424	18.509	163591.106	163430.463	-0.098