

Continuing Calibration Report for SequenceID : FE061325AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE054327.D

Aliphatic C9-C12	7747080.000	60.000	3.077	6.722	129118.000	125255.224	-3.084
Aliphatic C12-C16	5371336.000	40.000	6.723	10.169	134283.400	131151.686	-2.388
Aliphatic C16-C21	8195930.000	60.000	10.170	13.542	136598.833	133925.941	-1.996
Aliphatic C21-C28	10498600.000	80.000	13.543	17.208	131232.500	127068.779	-3.277
Aliphatic C28-C40	16119200.000	120.000	17.209	22.074	134326.667	128245.359	-4.742
Aliphatic EPH	47932146.000	360.000	3.077	22.074	133144.850	128755.230	-3.409