

Continuing Calibration Report for SequenceID : FE061325AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE054347.D

Aliphatic C9-C12	6972034.000	60.000	3.077	6.722	116200.567	125255.224	7.229
Aliphatic C12-C16	4901213.000	40.000	6.723	10.169	122530.325	131151.686	6.574
Aliphatic C16-C21	7534819.000	60.000	10.170	13.542	125580.317	133925.941	6.232
Aliphatic C21-C28	9651210.000	80.000	13.543	17.208	120640.125	127068.779	5.059
Aliphatic C28-C40	14974186.000	120.000	17.209	22.074	124784.883	128245.359	2.698
Aliphatic EPH	44033462.000	360.000	3.077	22.074	122315.172	128755.230	5.002