

Continuing Calibration Report for SequenceID : FE061325AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE054371.D

Aliphatic C9-C12	6809177.000	60.000	3.077	6.722	113486.283	125255.224	9.396
Aliphatic C12-C16	4837723.000	40.000	6.723	10.169	120943.075	131151.686	7.784
Aliphatic C16-C21	7604450.000	60.000	10.170	13.542	126740.833	133925.941	5.365
Aliphatic C21-C28	9877581.000	80.000	13.543	17.208	123469.763	127068.779	2.832
Aliphatic C28-C40	15340672.000	120.000	17.209	22.074	127838.933	128245.359	0.317
Aliphatic EPH	44469603.000	360.000	3.077	22.074	123526.675	128755.230	4.061