

Continuing Calibration Report for SequenceID : FF061625AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015992.D

Aromatic C10-C12	5856158.000	40.000	4.561	6.433	146403.950	127338.032	-14.973
Aromatic C12-C16	9096706.000	60.000	6.434	9.135	151611.767	131683.818	-15.133
Aromatic C16-C21	11861725.000	80.000	9.136	13.449	148271.563	129637.437	-14.374
Aromatic C21-C36	24382216.000	180.000	13.450	18.894	135456.756	121630.052	-11.368
Aromatic EPH	51196805.000	360.000	4.561	18.894	142213.347	125719.319	-13.120