

Continuing Calibration Report for SequenceID : FF061625AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF016002.D

Aromatic C10-C12	5523765.000	40.000	4.561	6.433	138094.125	127338.032	-8.447
Aromatic C12-C16	8586707.000	60.000	6.434	9.135	143111.783	131683.818	-8.678
Aromatic C16-C21	11181953.000	80.000	9.136	13.449	139774.413	129637.437	-7.819
Aromatic C21-C36	23205337.000	180.000	13.450	18.894	128918.539	121630.052	-5.992
Aromatic EPH	48497762.000	360.000	4.561	18.894	134716.006	125719.319	-7.156