

Continuing Calibration Report for SequenceID : FF061625AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FF016005.D

|                   |              |         |        |        |            |            |       |
|-------------------|--------------|---------|--------|--------|------------|------------|-------|
| Aliphatic C9-C12  | 7006386.000  | 60.000  | 3.267  | 6.898  | 116773.100 | 119334.527 | 2.146 |
| Aliphatic C12-C16 | 4858312.000  | 40.000  | 6.899  | 10.337 | 121457.800 | 124080.860 | 2.114 |
| Aliphatic C16-C21 | 7464918.000  | 60.000  | 10.338 | 13.706 | 124415.300 | 128733.722 | 3.355 |
| Aliphatic C21-C28 | 9717720.000  | 80.000  | 13.707 | 17.372 | 121471.500 | 127845.738 | 4.986 |
| Aliphatic C28-C40 | 11294553.000 | 120.000 | 17.373 | 22.341 | 94121.275  | 98908.616  | 4.840 |
| Aliphatic EPH     | 40341889.000 | 360.000 | 3.267  | 22.341 | 112060.803 | 116511.173 | 3.820 |