

Continuing Calibration Report for SequenceID : FF061625AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF016015.D

Aliphatic C9-C12	7114018.000	60.000	3.267	6.898	118566.967	119334.527	0.643
Aliphatic C12-C16	4916921.000	40.000	6.899	10.337	122923.025	124080.860	0.933
Aliphatic C16-C21	7516841.000	60.000	10.338	13.706	125280.683	128733.722	2.682
Aliphatic C21-C28	9789616.000	80.000	13.707	17.372	122370.200	127845.738	4.283
Aliphatic C28-C40	11567569.000	120.000	17.373	22.341	96396.408	98908.616	2.540
Aliphatic EPH	40904965.000	360.000	3.267	22.341	113624.903	116511.173	2.477