

Continuing Calibration Report for SequenceID : FG061625AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG016092.D

Aliphatic C9-C12	6887831.000	60.000	3.320	6.958	114797.183	109333.554	-4.997
Aliphatic C12-C16	4710401.000	40.000	6.959	10.415	117760.025	111824.611	-5.308
Aliphatic C16-C21	7219084.000	60.000	10.416	13.803	120318.067	114479.647	-5.100
Aliphatic C21-C28	9565184.000	80.000	13.804	17.486	119564.800	114390.192	-4.524
Aliphatic C28-C40	11444879.000	120.000	17.487	22.538	95373.992	91995.767	-3.672
Aliphatic EPH	39827379.000	360.000	3.320	22.538	110631.608	105812.455	-4.554