

Continuing Calibration Report for SequenceID : FG061625AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG016102.D

Aliphatic C9-C12	6539380.000	60.000	3.320	6.958	108989.667	109333.554	0.315
Aliphatic C12-C16	4464759.000	40.000	6.959	10.415	111618.975	111824.611	0.184
Aliphatic C16-C21	6846105.000	60.000	10.416	13.803	114101.750	114479.647	0.330
Aliphatic C21-C28	9124659.000	80.000	13.804	17.486	114058.238	114390.192	0.290
Aliphatic C28-C40	11126429.000	120.000	17.487	22.538	92720.242	91995.767	-0.788
Aliphatic EPH	38101332.000	360.000	3.320	22.538	105837.033	105812.455	-0.023