

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q2126 SAS No.: Q2126 SDG NO.: Q2126

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/18/2025 06/18/2025

Client Sample No.: CCAL02 Date Analyzed: 06/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS030765.D Time Analyzed: 20:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.667	9.568	9.768	759.000	712.500	6.5
2,4,5-TP(Silvex)	9.374	9.275	9.475	744.200	712.500	4.4
2,4-D	8.485	8.386	8.586	706.430	705.000	0.2
2,4-DB	10.246	10.147	10.347	806.080	712.500	13.1
2,4-DCAA	7.347	7.249	7.449	761.480	750.000	1.5
DICAMBA	7.538	7.439	7.639	744.220	705.000	5.6
DICHLORPROP	8.253	8.153	8.353	689.670	705.000	-2.2
Dinoseb	11.464	11.366	11.566	719.080	705.000	2.0

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q2126 SAS No.: Q2126 SDG NO.: Q2126

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/18/2025 06/18/2025

Client Sample No.: CCAL02 Date Analyzed: 06/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS030765.D Time Analyzed: 20:05

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.360	10.259	10.459	758.520	712.500	6.5
2,4,5-TP(Silvex)	9.934	9.833	10.033	745.360	712.500	4.6
2,4-D	9.028	8.927	9.127	710.320	705.000	0.8
2,4-DB	10.928	10.827	11.027	741.570	712.500	4.1
2,4-DCAA	7.771	7.670	7.870	770.560	750.000	2.7
DICAMBA	7.974	7.872	8.072	740.660	705.000	5.1
DICHLORPROP	8.693	8.591	8.791	699.550	705.000	-0.8
Dinoseb	11.313	11.211	11.411	720.610	705.000	2.2