

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q2126 SAS No.: Q2126 SDG NO.: Q2126

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/18/2025 06/18/2025

Client Sample No.: CCAL03 Date Analyzed: 06/20/2025

Lab Sample No.: HSTDCCC750 Data File : PS030788.D Time Analyzed: 19:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.663	9.568	9.768	784.130	712.500	10.1
2,4,5-TP(Silvex)	9.369	9.275	9.475	729.790	712.500	2.4
2,4-D	8.481	8.386	8.586	712.120	705.000	1.0
2,4-DB	10.240	10.147	10.347	831.140	712.500	16.7
2,4-DCAA	7.344	7.249	7.449	735.250	750.000	-2.0
DICAMBA	7.534	7.439	7.639	713.850	705.000	1.3
DICHLORPROP	8.248	8.153	8.353	664.490	705.000	-5.7
Dinoseb	11.459	11.366	11.566	724.080	705.000	2.7

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q2126 SAS No.: Q2126 SDG NO.: Q2126

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/18/2025 06/18/2025

Client Sample No.: CCAL03 Date Analyzed: 06/20/2025

Lab Sample No.: HSTDCCC750 Data File : PS030788.D Time Analyzed: 19:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.358	10.259	10.459	721.350	712.500	1.2
2,4,5-TP(Silvex)	9.932	9.833	10.033	709.270	712.500	-0.5
2,4-D	9.025	8.927	9.127	699.940	705.000	-0.7
2,4-DB	10.925	10.827	11.027	706.240	712.500	-0.9
2,4-DCAA	7.768	7.670	7.870	731.650	750.000	-2.4
DICAMBA	7.971	7.872	8.072	708.830	705.000	0.5
DICHLORPROP	8.690	8.591	8.791	671.760	705.000	-4.7
Dinoseb	11.310	11.211	11.411	694.590	705.000	-1.5