

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q2126 SAS No.: Q2126 SDG NO.: Q2126

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/18/2025 06/18/2025

Client Sample No.: CCAL04 Date Analyzed: 06/20/2025

Lab Sample No.: HSTDCCC750 Data File : PS030792.D Time Analyzed: 22:32

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.662	9.568	9.768	795.420	712.500	11.6
2,4,5-TP(Silvex)	9.368	9.275	9.475	752.730	712.500	5.6
2,4-D	8.480	8.386	8.586	722.630	705.000	2.5
2,4-DB	10.239	10.147	10.347	817.480	712.500	14.7
2,4-DCAA	7.344	7.249	7.449	750.580	750.000	0.1
DICAMBA	7.534	7.439	7.639	736.240	705.000	4.4
DICHLORPROP	8.248	8.153	8.353	684.990	705.000	-2.8
Dinoseb	11.457	11.366	11.566	751.060	705.000	6.5

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q2126 SAS No.: Q2126 SDG NO.: Q2126

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/18/2025 06/18/2025

Client Sample No.: CCAL04 Date Analyzed: 06/20/2025

Lab Sample No.: HSTDCCC750 Data File : PS030792.D Time Analyzed: 22:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.357	10.259	10.459	755.350	712.500	6.0
2,4,5-TP(Silvex)	9.931	9.833	10.033	742.250	712.500	4.2
2,4-D	9.026	8.927	9.127	707.450	705.000	0.3
2,4-DB	10.925	10.827	11.027	735.050	712.500	3.2
2,4-DCAA	7.768	7.670	7.870	766.230	750.000	2.2
DICAMBA	7.971	7.872	8.072	738.670	705.000	4.8
DICHLORPROP	8.690	8.591	8.791	703.380	705.000	-0.2
Dinoseb	11.310	11.211	11.411	734.870	705.000	4.2