

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS061825

Review By	Abdul	Review On	6/19/2025 8:25:06 AM
Supervise By	mohammad	Supervise On	6/20/2025 8:27:44 AM
SubDirectory	PS061825	HP Acquire Method	HP Processing Method PS061825 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC	PP24559		
Internal Standard/PEM			
ICV/I.BLK	PP24562		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS030737.D	18 Jun 2025 10:37		AR\AJ	Ok
2	I.BLK	I.BLK	PS030738.D	18 Jun 2025 11:01		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS030739.D	18 Jun 2025 11:25		AR\AJ	Ok,M
4	HSTDICC500	HSTDICC500	PS030740.D	18 Jun 2025 11:49		AR\AJ	Ok
5	HSTDICC750	HSTDICC750	PS030741.D	18 Jun 2025 12:13		AR\AJ	Ok,M
6	HSTDICC1000	HSTDICC1000	PS030742.D	18 Jun 2025 12:37		AR\AJ	Ok,M
7	HSTDICC1500	HSTDICC1500	PS030743.D	18 Jun 2025 13:29		AR\AJ	Ok,M
8	HSTDICV750	ICVPS061825	PS030744.D	18 Jun 2025 14:18		AR\AJ	Ok,M
9	I.BLK	I.BLK	PS030745.D	18 Jun 2025 14:44		AR\AJ	Ok
10	HSTDCCC750	HSTDCCC750	PS030746.D	18 Jun 2025 15:08		AR\AJ	Ok,M
11	Q2312-01MS	TP-1MS	PS030747.D	18 Jun 2025 17:07	some compounds fail for recovery	AR\AJ	Ok,M
12	Q2312-01MSD	TP-1MSD	PS030748.D	18 Jun 2025 17:31	some compounds fail for recovery & RPD	AR\AJ	Ok,M
13	I.BLK	I.BLK	PS030749.D	18 Jun 2025 17:55		AR\AJ	Ok
14	HSTDCCC750	HSTDCCC750	PS030750.D	18 Jun 2025 18:19		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS061925

Review By	Abdul	Review On	6/20/2025 8:24:30 AM
Supervise By	mohammad	Supervise On	6/20/2025 8:27:18 AM
SubDirectory	PS061925	HP Acquire Method	HP Processing Method PS061825 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24562
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS030751.D	19 Jun 2025 08:37		AR\AJ	Ok
2	I.BLK	I.BLK	PS030752.D	19 Jun 2025 09:01		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS030753.D	19 Jun 2025 09:25		AR\AJ	Ok,M
4	Q2266-01MS	WC-3MS	PS030754.D	19 Jun 2025 09:50	some compound recovery fail	AR\AJ	Ok,M
5	Q2266-01MSD	WC-3MSD	PS030755.D	19 Jun 2025 10:18	some compound recovery fail	AR\AJ	Ok,M
6	I.BLK	I.BLK	PS030756.D	19 Jun 2025 10:42		AR\AJ	Ok
7	HSTDCCC750	HSTDCCC750	PS030757.D	19 Jun 2025 11:06		AR\AJ	Ok,M
8	Q2126-01	LOD-MDL-SOIL-03-QT	PS030758.D	19 Jun 2025 17:16	Surrogate high in both column, LOD 100PPB	AR\AJ	Not Ok
9	Q2126-07	LOD-MDL-WATER-01-Q	PS030759.D	19 Jun 2025 17:40	Surrogate high in both column , LOD 100PPB	AR\AJ	Not Ok
10	Q2126-01	LOD-MDL-SOIL-03-QT	PS030760.D	19 Jun 2025 18:04	LOD 150 PPB	AR\AJ	Ok,M
11	Q2126-07	LOD-MDL-WATER-01-Q	PS030761.D	19 Jun 2025 18:29	LOD 150 PPB	AR\AJ	Ok,M
12	Q2126-02	LOQ-SOIL-02-QT2-202	PS030762.D	19 Jun 2025 18:53	LOQ 200 PPB	AR\AJ	Ok,M
13	Q2126-08	LOQ-WATER-02-QT2-2	PS030763.D	19 Jun 2025 19:17	LOQ 200 PPB	AR\AJ	Ok,M
14	I.BLK	I.BLK	PS030764.D	19 Jun 2025 19:41		AR\AJ	Ok
15	HSTDCCC750	HSTDCCC750	PS030765.D	19 Jun 2025 20:05		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS062025

Review By	Abdul	Review On	6/23/2025 9:05:06 AM
Supervise By	mohammad	Supervise On	6/24/2025 3:32:51 AM
SubDirectory	PS062025	HP Acquire Method	HP Processing Method PS061825 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24562
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS030766.D	20 Jun 2025 09:42		AR\AJ	Ok
2	I.BLK	I.BLK	PS030767.D	20 Jun 2025 10:06		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS030768.D	20 Jun 2025 10:30		AR\AJ	Ok,M
4	Q2365-01	BU-703-COMP-01	PS030769.D	20 Jun 2025 11:41		AR\AJ	Ok
5	Q2365-02	BU-703-COMP-02	PS030770.D	20 Jun 2025 12:05		AR\AJ	Ok,M
6	Q2365-03	BU-703-COMP-03	PS030771.D	20 Jun 2025 12:29		AR\AJ	Ok
7	Q2365-04	BU-703-COMP-04	PS030772.D	20 Jun 2025 12:53		AR\AJ	Ok,M
8	Q2365-05	BU-703-COMP-05	PS030773.D	20 Jun 2025 13:18		AR\AJ	Ok
9	Q2365-06	BU-703-COMP-06	PS030774.D	20 Jun 2025 13:42		AR\AJ	Ok,M
10	Q2365-07	BU-703-COMP-07	PS030775.D	20 Jun 2025 14:06		AR\AJ	Ok
11	Q2365-08	BU-703-COMP-08	PS030776.D	20 Jun 2025 14:30		AR\AJ	Ok
12	Q2365-09	BU-703-COMP-09	PS030777.D	20 Jun 2025 14:54		AR\AJ	Ok,M
13	I.BLK	I.BLK	PS030778.D	20 Jun 2025 15:18		AR\AJ	Ok
14	HSTDCCC750	HSTDCCC750	PS030779.D	20 Jun 2025 15:42		AR\AJ	Ok,M
15	PB168557BL	PB168557BL	PS030780.D	20 Jun 2025 16:06		AR\AJ	Ok
16	PB168557BS	PB168557BS	PS030781.D	20 Jun 2025 16:30	some compound recovery fail	AR\AJ	Not Ok
17	Q2347-01	TP-10	PS030782.D	20 Jun 2025 16:55		AR\AJ	Ok
18	Q2347-01MS	TP-10MS	PS030783.D	20 Jun 2025 17:19	some compound recovery fail	AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS062025

Review By	Abdul	Review On	6/23/2025 9:05:06 AM
Supervise By	mohammad	Supervise On	6/24/2025 3:32:51 AM
SubDirectory	PS062025	HP Acquire Method	HP Processing Method PS061825 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC	PP24559		
Internal Standard/PEM			
ICV/I.BLK	PP24562		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q2347-01MSD	TP-10MSD	PS030784.D	20 Jun 2025 17:43	some compound recovery fail	AR\AJ	Ok,M
20	Q2354-01	3321	PS030785.D	20 Jun 2025 18:07		AR\AJ	Ok
21	Q2355-01	3897	PS030786.D	20 Jun 2025 18:31		AR\AJ	Ok,M
22	I.BLK	I.BLK	PS030787.D	20 Jun 2025 18:55		AR\AJ	Ok
23	HSTDCCC750	HSTDCCC750	PS030788.D	20 Jun 2025 19:19	Comp#13 high in 1st column	AR\AJ	Ok,M
24	Q2126-01	LOD-MDL-SOIL-03-QT	PS030789.D	20 Jun 2025 19:43	LOD 100PPB	AR\AJ	Ok,M
25	Q2126-07	LOD-MDL-WATER-01-QT	PS030790.D	20 Jun 2025 20:07	LOD 100PPB	AR\AJ	Ok,M
26	I.BLK	I.BLK	PS030791.D	20 Jun 2025 20:32		AR\AJ	Ok
27	HSTDCCC750	HSTDCCC750	PS030792.D	20 Jun 2025 22:32		AR\AJ	Ok,M

M : Manual Integration