

Surrogate Summary

SDG No.: Q2126

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8151A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PS030738.D	PIBLK-PS030738.D	2,4-DCAA	1	500	414	83		61	136
		2,4-DCAA	2	500	487	97		61	136
I.BLK-PS030756.D	PIBLK-PS030756.D	2,4-DCAA	1	500	431	86		61	136
		2,4-DCAA	2	500	497	99		61	136
Q2126-01	LOD-MDL-SOIL-03-QT2-2025	2,4-DCAA	1	500	605	121		10	141
		2,4-DCAA	2	500	610	122		10	141
I.BLK-PS030764.D	PIBLK-PS030764.D	2,4-DCAA	1	500	453	91		61	136
		2,4-DCAA	2	500	514	103		61	136
I.BLK-PS030787.D	PIBLK-PS030787.D	2,4-DCAA	1	500	467	93		61	136
		2,4-DCAA	2	500	521	104		61	136
Q2126-01	LOD-MDL-SOIL-03-QT2-2025	2,4-DCAA	1	500	389	78		10	141
		2,4-DCAA	2	500	385	77		10	141
I.BLK-PS030791.D	PIBLK-PS030791.D	2,4-DCAA	1	500	475	95		61	136
		2,4-DCAA	2	500	523	105		61	136