

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q2126

Continuing Calib Date: 06/19/2025

Initial Calibration Date(s): 06/17/2025

06/17/2025

Continuing Calib Time: 23:06

Initial Calibration Time(s): 10:04

20:10

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.65       | 5.65      | 5.55      | 5.75  | 0.00       |
| Aroclor-1016-2       | (2) | 5.67       | 5.67      | 5.57      | 5.77  | 0.00       |
| Aroclor-1016-3       | (3) | 5.73       | 5.73      | 5.63      | 5.83  | 0.00       |
| Aroclor-1016-4       | (4) | 5.83       | 5.83      | 5.73      | 5.93  | 0.00       |
| Aroclor-1016-5       | (5) | 6.12       | 6.12      | 6.02      | 6.22  | 0.00       |
| Aroclor-1260-1       | (1) | 7.24       | 7.24      | 7.14      | 7.34  | 0.00       |
| Aroclor-1260-2       | (2) | 7.49       | 7.49      | 7.39      | 7.59  | 0.00       |
| Aroclor-1260-3       | (3) | 7.85       | 7.85      | 7.75      | 7.95  | 0.00       |
| Aroclor-1260-4       | (4) | 8.07       | 8.07      | 7.97      | 8.17  | 0.00       |
| Aroclor-1260-5       | (5) | 8.39       | 8.39      | 8.29      | 8.49  | 0.00       |
| Tetrachloro-m-xylene |     | 4.50       | 4.49      | 4.39      | 4.59  | 0.00       |
| Decachlorobiphenyl   |     | 10.19      | 10.19     | 10.09     | 10.29 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q2126

Continuing Calib Date: 06/19/2025

Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 23:06

Initial Calibration Time(s): 10:04 20:10

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.87       | 4.87      | 4.77      | 4.97 | 0.00       |
| Aroclor-1016-2       | (2) | 4.89       | 4.89      | 4.79      | 4.99 | 0.00       |
| Aroclor-1016-3       | (3) | 5.06       | 5.06      | 4.96      | 5.16 | 0.00       |
| Aroclor-1016-4       | (4) | 5.10       | 5.10      | 5.00      | 5.20 | 0.00       |
| Aroclor-1016-5       | (5) | 5.32       | 5.32      | 5.22      | 5.42 | 0.00       |
| Aroclor-1260-1       | (1) | 6.35       | 6.35      | 6.25      | 6.45 | 0.00       |
| Aroclor-1260-2       | (2) | 6.54       | 6.54      | 6.44      | 6.64 | 0.00       |
| Aroclor-1260-3       | (3) | 6.69       | 6.69      | 6.59      | 6.79 | 0.00       |
| Aroclor-1260-4       | (4) | 7.16       | 7.16      | 7.06      | 7.26 | 0.00       |
| Aroclor-1260-5       | (5) | 7.40       | 7.40      | 7.30      | 7.50 | 0.00       |
| Tetrachloro-m-xylene |     | 3.79       | 3.79      | 3.69      | 3.89 | 0.00       |
| Decachlorobiphenyl   |     | 8.80       | 8.80      | 8.70      | 8.90 | 0.00       |