

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q2126

Continuing Calib Date: 06/24/2025

Initial Calibration Date(s): 06/24/2025 06/24/2025

Continuing Calib Time: 18:35

Initial Calibration Time(s): 09:52 12:33

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.48	4.48	4.38	4.58	0.00
Aroclor-1016-2	(2)	4.50	4.50	4.40	4.60	0.00
Aroclor-1016-3	(3)	4.56	4.56	4.46	4.66	0.00
Aroclor-1016-4	(4)	4.65	4.65	4.55	4.75	0.00
Aroclor-1016-5	(5)	4.94	4.93	4.83	5.03	0.00
Aroclor-1260-1	(1)	6.02	6.02	5.92	6.12	0.00
Aroclor-1260-2	(2)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-3	(3)	6.63	6.63	6.53	6.73	0.00
Aroclor-1260-4	(4)	6.85	6.85	6.75	6.95	0.01
Aroclor-1260-5	(5)	7.15	7.15	7.05	7.25	0.00
Tetrachloro-m-xylene		3.40	3.40	3.30	3.50	0.00
Decachlorobiphenyl		8.50	8.50	8.40	8.60	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q2126

Continuing Calib Date: 06/24/2025

Initial Calibration Date(s): 06/24/2025 06/24/2025

Continuing Calib Time: 18:35

Initial Calibration Time(s): 09:52 12:33

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	3.78	3.78	3.68	3.88	0.00
Aroclor-1016-2	(2)	3.80	3.80	3.70	3.90	0.00
Aroclor-1016-3	(3)	3.96	3.96	3.86	4.06	0.00
Aroclor-1016-4	(4)	4.01	4.01	3.91	4.11	0.00
Aroclor-1016-5	(5)	4.20	4.20	4.10	4.30	0.00
Aroclor-1260-1	(1)	5.20	5.20	5.10	5.30	0.00
Aroclor-1260-2	(2)	5.39	5.39	5.29	5.49	0.00
Aroclor-1260-3	(3)	5.53	5.53	5.43	5.63	0.00
Aroclor-1260-4	(4)	5.99	5.99	5.89	6.09	0.00
Aroclor-1260-5	(5)	6.23	6.23	6.13	6.33	0.00
Tetrachloro-m-xylene		2.78	2.78	2.68	2.88	0.00
Decachlorobiphenyl		7.54	7.54	7.44	7.64	0.00