

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2126
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/24/2025 06/24/2025

Client Sample No.: CCAL02 **Date Analyzed:** 06/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ070431.D **Time Analyzed:** 02:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.483	4.384	4.584	507.760	500.000	1.6
Aroclor-1016-2	4.502	4.403	4.603	499.790	500.000	0.0
Aroclor-1016-3	4.560	4.461	4.661	484.220	500.000	-3.2
Aroclor-1016-4	4.650	4.550	4.750	489.390	500.000	-2.1
Aroclor-1016-5	4.934	4.834	5.034	476.550	500.000	-4.7
Aroclor-1260-1	6.024	5.924	6.124	488.570	500.000	-2.3
Aroclor-1260-2	6.280	6.180	6.380	511.390	500.000	2.3
Aroclor-1260-3	6.631	6.531	6.731	485.010	500.000	-3.0
Aroclor-1260-4	6.844	6.745	6.945	501.440	500.000	0.3
Aroclor-1260-5	7.150	7.049	7.249	531.040	500.000	6.2
Decachlorobiphenyl	8.500	8.402	8.602	51.450	50.000	2.9
Tetrachloro-m-xylene	3.404	3.304	3.504	49.040	50.000	-1.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2126
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/24/2025 06/24/2025

Client Sample No.: CCAL02 **Date Analyzed:** 06/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ070431.D **Time Analyzed:** 02:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.783	3.684	3.884	525.690	500.000	5.1
Aroclor-1016-2	3.799	3.699	3.899	493.500	500.000	-1.3
Aroclor-1016-3	3.959	3.860	4.060	510.430	500.000	2.1
Aroclor-1016-4	4.007	3.908	4.108	473.890	500.000	-5.2
Aroclor-1016-5	4.203	4.103	4.303	481.570	500.000	-3.7
Aroclor-1260-1	5.196	5.097	5.297	472.910	500.000	-5.4
Aroclor-1260-2	5.385	5.286	5.486	500.670	500.000	0.1
Aroclor-1260-3	5.526	5.428	5.628	490.860	500.000	-1.8
Aroclor-1260-4	5.987	5.888	6.088	489.490	500.000	-2.1
Aroclor-1260-5	6.231	6.132	6.332	514.510	500.000	2.9
Decachlorobiphenyl	7.543	7.443	7.643	50.220	50.000	0.4
Tetrachloro-m-xylene	2.782	2.682	2.882	47.270	50.000	-5.5