

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: ALLI03
 ProjectID: LOD-LOQ-MDL Study
 Lab Code: CHEM Case No.: Q2126 SAS No.: Q2126 SDG No.: Q2126

Calibration Sequence : FF052125		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	204184917	120109	FF015874.D
850	105072733	123615	FF015875.D
340	41918129	123289	FF015876.D
170	23592968	138782	FF015877.D
85	10584775	124527	FF015878.D
AVG RF : 126064		% RSD : 5.792	AVG RT : 15.0206