



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q2126

Continuing Calib Date: 08/07/2025

Initial Calibration Date(s): 08/07/202508/07/2025

Continuing Calib Time: 09:58

Initial Calibration Time(s): 08:4209:21

GC Column: ZB-MR1ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DBCP	5.96	5.96	5.86	6.06	0.00
EDB	2.61	2.61	2.51	2.71	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q2126

Continuing Calib Date: 08/07/2025

Initial Calibration Date(s): 08/07/2025 08/07/2025

Continuing Calib Time: 09:58

Initial Calibration Time(s): 08:42 09:21

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DBCP	4.99	4.99	4.89	5.09	0.00
EDB	1.98	1.98	1.88	2.08	0.00