

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2126
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/07/2025
Lab Sample No.: M8011.504.1 0.1 PF **Data File :** PQ070736.D **Time Analyzed:** 13:41

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
DBCP	5.959	5.857	6.057	0.100	0.100	0.0
EDB	2.608	2.507	2.707	0.100	0.100	0.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2126
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/07/2025
Lab Sample No.: M8011.504.1 0.1 PF **Data File :** PQ070736.D **Time Analyzed:** 13:41

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
DBCP	4.988	4.888	5.088	0.090	0.100	0.0
EDB	1.975	1.875	2.075	0.110	0.100	0.0