

Instrument ID: IC-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135928

Review By	rubina	Review On	5/30/2025 11:40:18 AM
Supervise By	Iwona	Supervise On	5/30/2025 12:35:36 PM
SubDirectory	LB135928	Test	Anions
STD. NAME	STD REF.#		
ICAL Standard	WP113186,WP113187,WP113188,WP113189,WP113190,WP113191,WP113192		
ICV Standard	WP113193		
CCV Standard	WP113251		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP113252		
Chk Standard	WP113194,WP113195,WP113252,WP113254,WP113255		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	STD1	STD1	CAL1	05/22/25 11:09	All standards, samples, and	NF/IZ	OK
2	STD2	STD2	CAL2	05/22/25 11:30	QC are filtered through	NF/IZ	OK
3	STD3	STD3	CAL3	05/22/25 11:52	0.45um, filter lot W3160	NF/IZ	OK
4	STD4	STD4	CAL4	05/22/25 12:13		NF/IZ	OK
5	STD5	STD5	CAL5	05/22/25 12:35		NF/IZ	OK
6	STD6	STD6	CAL6	05/22/25 12:56		NF/IZ	OK
7	STD7	STD7	CAL7	05/22/25 13:17		NF/IZ	OK
8	ICV1	ICV1	ICV	05/22/25 13:39		NF/IZ	OK
9	ICB1	ICB1	ICB	05/22/25 14:22		NF/IZ	OK
10	CCV1	CCV1	CCV	05/28/25 11:29		NF/IZ	OK
11	CCB1	CCB1	CCB	05/28/25 11:51		NF/IZ	OK
12	LB135928BLW	LB135928BLW	MB	05/28/25 12:12		NF/IZ	OK
13	LB135928BSW	LB135928BSW	LCS	05/28/25 13:29		NF/IZ	OK
14	Q2126-08	LOQ-WATER-02-QT2	LOQ	05/28/25 13:51		NF/IZ	OK
15	Q2126-07	LOD-MDL-WATER-01	SAM	05/28/25 14:12		NF/IZ	OK
16	Q2126-07RE	LOD-MDL-WATER-01	SAM	05/28/25 14:34		NF/IZ	OK
17	Q2133-01	MW1	SAM	05/28/25 14:55	Cl is high	NF/IZ	Dilution
18	Q2133-02	MW5	SAM	05/28/25 15:17	Cl and SO4 is high	NF/IZ	Dilution

Instrument ID: IC-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135928

Review By	rubina	Review On	5/30/2025 11:40:18 AM
Supervise By	Iwona	Supervise On	5/30/2025 12:35:36 PM
SubDirectory	LB135928	Test	Anions
STD. NAME	STD REF.#		
ICAL Standard	WP113186,WP113187,WP113188,WP113189,WP113190,WP113191,WP113192		
ICV Standard	WP113193		
CCV Standard	WP113251		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP113252		
Chk Standard	WP113194,WP113195,WP113252,WP113254,WP113255		

19	Q2133-03	SUPPLY	SAM	05/28/25 15:38	Cl is high	NF/IZ	Dilution
20	Q2133-03MS	SUPPLYMS	MS	05/28/25 16:00	9.5ml of sample, 0.5mL W3092	NF/IZ	OK
21	Q2133-03MSD	SUPPLYMSD	MSD	05/28/25 16:21	9.5ml of sample, 0.5mL W3092	NF/IZ	OK
22	CCV2	CCV2	CCV	05/28/25 16:43		NF/IZ	OK
23	CCB2	CCB2	CCB	05/28/25 17:05		NF/IZ	OK
24	Q2133-01DL	MW1DL	SAM	05/28/25 17:26	100X For Cl	NF/IZ	Confirms
25	Q2133-02DL	MW5DL	SAM	05/28/25 18:09	100X for CL and SO4, still Cl is high	NF/IZ	Dilution
26	Q2133-02DL2	MW5DL2	SAM	05/28/25 18:52	1000X for Cl	NF/IZ	Confirms
27	Q2133-03DL	SUPPLYDL	SAM	05/28/25 19:35	100X for Cl	NF/IZ	Confirms
28	CCV3	CCV3	CCV	05/28/25 19:56		NF/IZ	OK
29	CCB3	CCB3	CCB	05/28/25 20:18		NF/IZ	OK