

Laboratory Control Sample Summary

Client: G Environmental

SDG No.: Q2133

Project: Banker

Run No.: LB135928

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135928BSW							
Bromide	mg/L	10	10.7		107	1	90-110	05/28/2025
Chloride	mg/L	3	3.30		110	1	90-110	05/28/2025
Fluoride	mg/L	2	2.20		110	1	90-110	05/28/2025
Nitrite	mg/L	3	3.20		107	1	90-110	05/28/2025
Nitrate	mg/L	2.5	2.70		108	1	90-110	05/28/2025
Sulfate	mg/L	15	15.9		106	1	90-110	05/28/2025
Orthophosphate as P	mg/L	5	5.40		108	1	90-110	05/28/2025