



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental **SDG No.:** Q2134
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2134 **SAS No.:** Q2134
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Iron	106	100	106	65 - 135	P	06/02/2025	13:08	LB135981
	Manganese	20.1	20.0	101	65 - 135	P	06/02/2025	13:08	LB135981
	Sodium	1690	2000	85	65 - 135	P	06/02/2025	13:08	LB135981
CRI01	Iron	119	100	119	65 - 135	P	06/03/2025	14:36	LB135993
	Manganese	19.4	20.0	97	65 - 135	P	06/03/2025	14:36	LB135993
	Sodium	1810	2000	90	65 - 135	P	06/03/2025	14:36	LB135993