

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>G Environmental</u>	level: <u>low</u>	sdg no.: <u>Q2134</u>
contract: <u>GENV01</u>	lab code: <u>CHEM</u>	case no.: <u>Q2134</u> sas no.: <u>Q2134</u>
matrix: <u>Water</u>	sample id: <u>Q2133-03</u>	client id: <u>SUPPLYMS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2133-03MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	75 - 125	1200		51.1		1500	77		P
Manganese	ug/L	75 - 125	81.3		13.7		100	68	N	P
Sodium	ug/L	75 - 125	46700		57500		1500	-722		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>G Environmental</u>	level: <u>low</u>	sdg no.: <u>Q2134</u>
contract: <u>GENV01</u>	lab code: <u>CHEM</u>	case no.: <u>Q2134</u> sas no.: <u>Q2134</u>
matrix: <u>Water</u>	sample id: <u>Q2133-03</u>	client id: <u>SUPPLYMSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2133-03MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	75 - 125	1430		51.1		1500	92		P
Manganese	ug/L	75 - 125	88.4		13.7		100	75		P
Sodium	ug/L	75 - 125	57400		57500		1500	-5		P