

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF052025

Review By	Rahul	Review On	5/21/2025 2:52:20 PM
Supervise By	Jagrut	Supervise On	5/21/2025 4:44:06 PM
SubDirectory	BF052025	HP Acquire Method	BNA_F
		HP Processing Method	bf052025

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	SP6757 SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6787 S12665,10ul/1000ul sample SP6770

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142465.D	20 May 2025 11:13		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142466.D	20 May 2025 11:41	Fresh Calibration Required	RC/JU	Not Ok
3	SSTDICC2.5	SSTDICC2.5	BF142467.D	20 May 2025 12:10		RC/JU	Ok
4	SSTDICC005	SSTDICC005	BF142468.D	20 May 2025 12:38	Compound #32,54,85 removed from 5ppm	RC/JU	Ok
5	SSTDICC010	SSTDICC010	BF142469.D	20 May 2025 13:07		RC/JU	Ok,M
6	SSTDICC020	SSTDICC020	BF142470.D	20 May 2025 13:36		RC/JU	Ok
7	SSTDICCC040	SSTDICCC040	BF142471.D	20 May 2025 14:05	This calibration is good for both the methods, 8270E DOD and 625.1.	RC/JU	Ok
8	SSTDICC050	SSTDICC050	BF142472.D	20 May 2025 14:34		RC/JU	Ok
9	SSTDICC060	SSTDICC060	BF142473.D	20 May 2025 15:03		RC/JU	Ok
10	SSTDICC080	SSTDICC080	BF142474.D	20 May 2025 15:31		RC/JU	Ok
11	SSTDICV040	ICVBF052025	BF142475.D	20 May 2025 16:31		RC/JU	Ok
12	PB168067TB	PB168067TB	BF142476.D	20 May 2025 17:29		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF060225

Review By	Rahul	Review On	6/3/2025 1:30:43 PM
Supervise By	Jagrut	Supervise On	6/3/2025 5:25:27 PM
SubDirectory	BF060225	HP Acquire Method	BNA_F
		HP Processing Method	bf052025

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791
CCC	SP6787
Internal Standard/PEM	S12667,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142589.D	02 Jun 2025 11:01		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142590.D	02 Jun 2025 11:30		RC/JU	Ok
3	PB168219BL	PB168219BL	BF142591.D	02 Jun 2025 11:59		RC/JU	Ok
4	PB168219BS	PB168219BS	BF142592.D	02 Jun 2025 12:28		RC/JU	Ok,M
5	PB168190TB	PB168190TB	BF142593.D	02 Jun 2025 12:57		RC/JU	Ok
6	PB168235BL	PB168235BL	BF142594.D	02 Jun 2025 13:26		RC/JU	Ok
7	PB168235BS	PB168235BS	BF142595.D	02 Jun 2025 13:54		RC/JU	Ok,M
8	PB168235BSD	PB168235BSD	BF142596.D	02 Jun 2025 14:23		RC/JU	Ok,M
9	Q2134-01	MW10	BF142597.D	02 Jun 2025 15:05		RC/JU	Ok
10	Q2150-01DL	TP-44DL	BF142598.D	02 Jun 2025 15:34		RC/JU	Ok,M
11	Q2149-01	FILTER-CAKE	BF142599.D	02 Jun 2025 16:03	Need 5X Dilution	RC/JU	Dilution
12	Q2149-01DL	FILTER-CAKEDL	BF142600.D	02 Jun 2025 17:11		RC/JU	Ok

M : Manual Integration