

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q2134 SAS No.: Q2134 SDG No.: Q2134

Instrument ID: MSVOA_N Calibration Date/Time: 05/28/2025 10:58

Lab File ID: VN086790.D Init. Calib. Date(s): 05/16/2025 05/16/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 16:31 18:32

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
1,2-Dichloroethane-d4	0.691	0.651		-5.79	20
Dibromofluoromethane	0.298	0.284		-4.7	20
Toluene-d8	1.234	1.176		-4.7	20
4-Bromofluorobenzene	0.436	0.436		0	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.