

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2134 SAS No.: Q2134 SDG No.: Q2134
 Instrument ID: MSVOA_N Calibration Date(s): 05/16/2025 05/16/2025
 Heated Purge: (Y/N) N Calibration Time(s): 16:31 18:32
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF001 = VN086669.D		RRF005 = VN086670.D		RRF010 = VN086671.D		
		RRF020 = VN086672.D		RRF050 = VN086673.D		RRF100 = VN086674.D		
COMPOUND	RRF001	RRF005	RRF010	RRF020	RRF050	RRF100	$\overline{\text{RRF}}$	% RSD
1,2-Dichloroethane-d4		0.718	0.658	0.687	0.693	0.698	0.691	3.2
Dibromofluoromethane		0.300	0.289	0.291	0.298	0.311	0.298	3
Toluene-d8		1.203	1.159	1.233	1.250	1.325	1.234	5
4-Bromofluorobenzene		0.412	0.403	0.433	0.450	0.481	0.436	7.2

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2134 SAS No.: Q2134 SDG No.: Q2134
 Instrument ID: MSVOA_X Calibration Date(s): 05/05/2025 05/05/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:35 16:27
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF020 = VX046041.D		RRF050 = VX046042.D		RRF100 = VX046043.D		
		RRF150 = VX046044.D		RRF005 = VX046046.D		RRF001 = VX046047.D		
COMPOUND	RRF020	RRF050	RRF100	RRF150	RRF005	RRF001	RRF	% RSD
1,2-Dichloroethane-d4	0.953	0.910	0.930	0.932	0.935		0.932	1.6
Dibromofluoromethane	0.359	0.355	0.364	0.368	0.354		0.360	1.7
Toluene-d8	1.246	1.223	1.266	1.275	1.221		1.246	2
4-Bromofluorobenzene	0.455	0.470	0.500	0.500	0.464		0.478	4.4

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.