

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2134 SAS No.: Q2134 SDG NO.: Q2134
 Lab File ID: VN086790.D Date Analyzed: 05/28/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:58
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	194902	8.22	356866	9.09	328898	11.86
UPPER LIMIT	389804	8.718	713732	9.594	657796	12.359
LOWER LIMIT	97451	7.718	178433	8.594	164449	11.359
EPA SAMPLE NO.						
MW10	168502	8.22	319448	9.10	306754	11.87
VN0528WBL01	190621	8.22	373140	9.10	366675	11.87
VN0528WBS01	204156	8.22	364096	9.10	323600	11.87
VN0528WBSD01	194862	8.22	348503	9.10	308495	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.