

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q2134 SAS No.: Q2134 SDG NO.: Q2134
Lab File ID: VX046390.D Date Analyzed: 05/29/2025
Instrument ID: MSVOA_X Time Analyzed: 11:27
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	82744	5.54	141651	6.75	125856	10.05
UPPER LIMIT	165488	6.038	283302	7.251	251712	10.549
LOWER LIMIT	41372	5.038	70825.5	6.251	62928	9.549
EPA SAMPLE NO.						
MW10	60016	5.55	114311	6.76	110287	10.05
VX0529WBL01	60328	5.54	121532	6.76	115333	10.06
VX0529WBS01	84888	5.54	151471	6.76	132494	10.05
VX0529WBSD01	81952	5.54	144475	6.76	129368	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.