

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2134 SAS No.: Q2134 SDG NO.: Q2134
 Lab File ID: VN086790.D Date Analyzed: 05/28/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:58
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	156869	13.788				
UPPER LIMIT	313738	14.288				
LOWER LIMIT	78434.5	13.288				
EPA SAMPLE NO.						
MW10	144822	13.79				
VN0528WBL01	154289	13.79				
VN0528WBS01	153229	13.79				
VN0528WBSD01	143771	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.