

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2134 SAS No.: Q2134 SDG NO.: Q2134
 Lab File ID: VX046390.D Date Analyzed: 05/29/2025
 Instrument ID: MSVOA_X Time Analyzed: 11:27
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	61906	12.018				
UPPER LIMIT	123812	12.518				
LOWER LIMIT	30953	11.518				
EPA SAMPLE NO.						
MW10	46578	12.02				
VX0529WBL01	50276	12.02				
VX0529WBS01	62151	12.02				
VX0529WBSD01	61624	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.