

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN052825

Review By	John Carlone	Review On	5/29/2025 9:58:36 AM
Supervise By	Maresh Dadoda	Supervise On	5/29/2025 2:01:51 PM
SubDirectory	VN052825	HP Acquire Method	HP Processing Method 82N051625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134041		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134042,VP134043		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN086789.D	28 May 2025 08:34	JC\MD	Ok
2	VSTDCCC050	VN086790.D	28 May 2025 10:58	JC\MD	Ok,M
3	VN0528MBL01	VN086791.D	28 May 2025 11:35	JC\MD	Ok
4	VN0528WBL01	VN086792.D	28 May 2025 11:59	JC\MD	Ok
5	VN0528WBS01	VN086793.D	28 May 2025 12:26	JC\MD	Ok,M
6	VN0528WBSD01	VN086794.D	28 May 2025 12:59	JC\MD	Ok,M
7	Q2134-01	VN086795.D	28 May 2025 13:23	JC\MD	ReRun
8	IBLK	VN086796.D	28 May 2025 13:47	JC\MD	Ok
9	IBLK	VN086797.D	28 May 2025 14:12	JC\MD	Ok
10	Q2126-09 0.2PPB	VN086798.D	28 May 2025 14:36	JC\MD	ReRun
11	Q2126-09 0.5PPB	VN086799.D	28 May 2025 15:09	JC\MD	ReRun
12	Q2126-09 0.75PPB	VN086800.D	28 May 2025 15:33	JC\MD	ReRun
13	Q2126-09 2.5PPB	VN086801.D	28 May 2025 15:57	JC\MD	ReRun
14	VSTDCCC050	VN086802.D	28 May 2025 16:34	JC\MD	Ok,M

M : Manual Integration