

### CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q2135 SAS No.: Q2135 SDG NO.: Q2135

Continuing Calib Date: 06/02/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 14:57 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.65       | 5.65      | 5.55      | 5.75  | 0.00       |
| Aroclor-1016-2       | (2) | 5.67       | 5.67      | 5.57      | 5.77  | 0.00       |
| Aroclor-1016-3       | (3) | 5.74       | 5.74      | 5.64      | 5.84  | 0.00       |
| Aroclor-1016-4       | (4) | 5.83       | 5.83      | 5.73      | 5.93  | 0.00       |
| Aroclor-1016-5       | (5) | 6.13       | 6.13      | 6.03      | 6.23  | 0.00       |
| Aroclor-1260-1       | (1) | 7.24       | 7.25      | 7.15      | 7.35  | 0.01       |
| Aroclor-1260-2       | (2) | 7.50       | 7.50      | 7.40      | 7.60  | 0.00       |
| Aroclor-1260-3       | (3) | 7.86       | 7.86      | 7.76      | 7.96  | 0.00       |
| Aroclor-1260-4       | (4) | 8.08       | 8.09      | 7.99      | 8.19  | 0.01       |
| Aroclor-1260-5       | (5) | 8.40       | 8.40      | 8.30      | 8.50  | 0.00       |
| Tetrachloro-m-xylene |     | 4.50       | 4.50      | 4.40      | 4.60  | 0.00       |
| Decachlorobiphenyl   |     | 10.20      | 10.21     | 10.11     | 10.31 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q2135 SAS No.: Q2135 SDG NO.: Q2135

Continuing Calib Date: 06/02/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 14:57 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.88       | 4.88      | 4.78      | 4.98 | 0.00       |
| Aroclor-1016-2       | (2) | 4.90       | 4.90      | 4.80      | 5.00 | 0.01       |
| Aroclor-1016-3       | (3) | 5.07       | 5.07      | 4.97      | 5.17 | 0.00       |
| Aroclor-1016-4       | (4) | 5.11       | 5.12      | 5.02      | 5.22 | 0.01       |
| Aroclor-1016-5       | (5) | 5.33       | 5.33      | 5.23      | 5.43 | 0.00       |
| Aroclor-1260-1       | (1) | 6.36       | 6.36      | 6.26      | 6.46 | 0.00       |
| Aroclor-1260-2       | (2) | 6.55       | 6.55      | 6.45      | 6.65 | 0.00       |
| Aroclor-1260-3       | (3) | 6.70       | 6.71      | 6.61      | 6.81 | 0.01       |
| Aroclor-1260-4       | (4) | 7.17       | 7.18      | 7.08      | 7.28 | 0.01       |
| Aroclor-1260-5       | (5) | 7.41       | 7.42      | 7.32      | 7.52 | 0.01       |
| Tetrachloro-m-xylene |     | 3.80       | 3.79      | 3.69      | 3.89 | -0.01      |
| Decachlorobiphenyl   |     | 8.81       | 8.82      | 8.72      | 8.92 | 0.01       |