

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q2135 SAS No.: Q2135 SDG NO.: Q2135

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL01 Date Analyzed: 05/30/2025

Lab Sample No.: AR1660CCC500 Data File : PO111355.D Time Analyzed: 12:39

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.770	4.673	4.873	521.980	500.000	4.4
Aroclor-1016-2	4.790	4.692	4.892	520.880	500.000	4.2
Aroclor-1016-3	4.845	4.749	4.949	513.620	500.000	2.7
Aroclor-1016-4	4.966	4.869	5.069	545.820	500.000	9.2
Aroclor-1016-5	5.223	5.126	5.326	506.390	500.000	1.3
Aroclor-1260-1	6.263	6.166	6.366	512.730	500.000	2.5
Aroclor-1260-2	6.453	6.355	6.555	523.300	500.000	4.7
Aroclor-1260-3	6.820	6.722	6.922	501.720	500.000	0.3
Aroclor-1260-4	7.080	6.983	7.183	490.320	500.000	-1.9
Aroclor-1260-5	7.323	7.225	7.425	484.970	500.000	-3.0
Decachlorobiphenyl	8.723	8.624	8.824	44.220	50.000	-11.6
Tetrachloro-m-xylene	3.681	3.582	3.782	49.040	50.000	-1.9

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q2135 SAS No.: Q2135 SDG NO.: Q2135

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL01 Date Analyzed: 05/30/2025

Lab Sample No.: AR1660CCC500 Data File : PO111355.D Time Analyzed: 12:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.756	4.658	4.858	463.180	500.000	-7.4
Aroclor-1016-2	4.775	4.677	4.877	463.220	500.000	-7.4
Aroclor-1016-3	4.950	4.852	5.052	468.920	500.000	-6.2
Aroclor-1016-4	4.992	4.894	5.094	466.900	500.000	-6.6
Aroclor-1016-5	5.205	5.107	5.307	461.280	500.000	-7.7
Aroclor-1260-1	6.236	6.138	6.338	455.480	500.000	-8.9
Aroclor-1260-2	6.424	6.326	6.526	447.610	500.000	-10.5
Aroclor-1260-3	6.576	6.479	6.679	422.070	500.000	-15.6
Aroclor-1260-4	7.046	6.949	7.149	438.160	500.000	-12.4
Aroclor-1260-5	7.288	7.191	7.391	429.910	500.000	-14.0
Decachlorobiphenyl	8.672	8.574	8.774	42.230	50.000	-15.5
Tetrachloro-m-xylene	3.677	3.579	3.779	47.250	50.000	-5.5