

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q2135 SAS No.: Q2135 SDG NO.: Q2135

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL02 Date Analyzed: 05/30/2025

Lab Sample No.: AR1660CCC500 Data File : PO111368.D Time Analyzed: 19:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.768	4.673	4.873	525.180	500.000	5.0
Aroclor-1016-2	4.787	4.692	4.892	524.270	500.000	4.9
Aroclor-1016-3	4.843	4.749	4.949	527.720	500.000	5.5
Aroclor-1016-4	4.964	4.869	5.069	550.960	500.000	10.2
Aroclor-1016-5	5.221	5.126	5.326	494.550	500.000	-1.1
Aroclor-1260-1	6.261	6.166	6.366	521.430	500.000	4.3
Aroclor-1260-2	6.450	6.355	6.555	524.220	500.000	4.8
Aroclor-1260-3	6.818	6.722	6.922	498.610	500.000	-0.3
Aroclor-1260-4	7.078	6.983	7.183	489.690	500.000	-2.1
Aroclor-1260-5	7.320	7.225	7.425	485.400	500.000	-2.9
Decachlorobiphenyl	8.720	8.624	8.824	43.890	50.000	-12.2
Tetrachloro-m-xylene	3.679	3.582	3.782	49.430	50.000	-1.1

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q2135 SAS No.: Q2135 SDG NO.: Q2135

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL02 Date Analyzed: 05/30/2025

Lab Sample No.: AR1660CCC500 Data File : PO111368.D Time Analyzed: 19:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.755	4.658	4.858	474.220	500.000	-5.2
Aroclor-1016-2	4.774	4.677	4.877	473.820	500.000	-5.2
Aroclor-1016-3	4.949	4.852	5.052	474.200	500.000	-5.2
Aroclor-1016-4	4.992	4.894	5.094	471.970	500.000	-5.6
Aroclor-1016-5	5.204	5.107	5.307	465.150	500.000	-7.0
Aroclor-1260-1	6.235	6.138	6.338	456.040	500.000	-8.8
Aroclor-1260-2	6.424	6.326	6.526	448.090	500.000	-10.4
Aroclor-1260-3	6.575	6.479	6.679	420.090	500.000	-16.0
Aroclor-1260-4	7.047	6.949	7.149	431.890	500.000	-13.6
Aroclor-1260-5	7.288	7.191	7.391	420.860	500.000	-15.8
Decachlorobiphenyl	8.672	8.574	8.774	42.050	50.000	-15.9
Tetrachloro-m-xylene	3.677	3.579	3.779	48.450	50.000	-3.1