

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q2135 SAS No.: Q2135 SDG NO.: Q2135

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 06/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP072497.D Time Analyzed: 08:56

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.651	5.552	5.752	423.220	500.000	-15.4
Aroclor-1016-2	5.673	5.574	5.774	458.410	500.000	-8.3
Aroclor-1016-3	5.735	5.636	5.836	457.770	500.000	-8.4
Aroclor-1016-4	5.832	5.734	5.934	470.390	500.000	-5.9
Aroclor-1016-5	6.125	6.027	6.227	458.940	500.000	-8.2
Aroclor-1260-1	7.242	7.147	7.347	435.960	500.000	-12.8
Aroclor-1260-2	7.496	7.401	7.601	438.020	500.000	-12.4
Aroclor-1260-3	7.854	7.760	7.960	458.230	500.000	-8.4
Aroclor-1260-4	8.078	7.985	8.185	448.300	500.000	-10.3
Aroclor-1260-5	8.397	8.303	8.503	468.660	500.000	-6.3
Decachlorobiphenyl	10.202	10.113	10.313	46.930	50.000	-6.1
Tetrachloro-m-xylene	4.499	4.396	4.596	47.890	50.000	-4.2

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q2135 SAS No.: Q2135 SDG NO.: Q2135

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 06/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP072497.D Time Analyzed: 08:56

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.875	4.777	4.977	478.640	500.000	-4.3
Aroclor-1016-2	4.893	4.795	4.995	489.940	500.000	-2.0
Aroclor-1016-3	5.070	4.973	5.173	495.190	500.000	-1.0
Aroclor-1016-4	5.111	5.015	5.215	505.450	500.000	1.1
Aroclor-1016-5	5.325	5.229	5.429	488.380	500.000	-2.3
Aroclor-1260-1	6.358	6.264	6.464	500.130	500.000	0.0
Aroclor-1260-2	6.546	6.453	6.653	476.210	500.000	-4.8
Aroclor-1260-3	6.698	6.606	6.806	497.820	500.000	-0.4
Aroclor-1260-4	7.169	7.077	7.277	496.800	500.000	-0.6
Aroclor-1260-5	7.411	7.319	7.519	487.500	500.000	-2.5
Decachlorobiphenyl	8.810	8.723	8.923	48.310	50.000	-3.4
Tetrachloro-m-xylene	3.793	3.691	3.891	49.540	50.000	-0.9