

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q2135 SAS No.: Q2135 SDG NO.: Q2135

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL04 Date Analyzed: 06/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP072512.D Time Analyzed: 14:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.653	5.552	5.752	414.600	500.000	-17.1
Aroclor-1016-2	5.674	5.574	5.774	460.370	500.000	-7.9
Aroclor-1016-3	5.737	5.636	5.836	451.410	500.000	-9.7
Aroclor-1016-4	5.834	5.734	5.934	467.550	500.000	-6.5
Aroclor-1016-5	6.126	6.027	6.227	457.220	500.000	-8.6
Aroclor-1260-1	7.244	7.147	7.347	451.300	500.000	-9.7
Aroclor-1260-2	7.498	7.401	7.601	452.650	500.000	-9.5
Aroclor-1260-3	7.856	7.760	7.960	475.260	500.000	-4.9
Aroclor-1260-4	8.080	7.985	8.185	464.740	500.000	-7.1
Aroclor-1260-5	8.399	8.303	8.503	474.670	500.000	-5.1
Decachlorobiphenyl	10.203	10.113	10.313	48.520	50.000	-3.0
Tetrachloro-m-xylene	4.501	4.396	4.596	47.000	50.000	-6.0

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q2135 SAS No.: Q2135 SDG NO.: Q2135

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL04 Date Analyzed: 06/02/2025

Lab Sample No.: AR1660CCC500 Data File : PP072512.D Time Analyzed: 14:57

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.877	4.777	4.977	486.380	500.000	-2.7
Aroclor-1016-2	4.895	4.795	4.995	497.120	500.000	-0.6
Aroclor-1016-3	5.071	4.973	5.173	506.770	500.000	1.4
Aroclor-1016-4	5.114	5.015	5.215	516.720	500.000	3.3
Aroclor-1016-5	5.328	5.229	5.429	504.200	500.000	0.8
Aroclor-1260-1	6.360	6.264	6.464	508.570	500.000	1.7
Aroclor-1260-2	6.548	6.453	6.653	480.070	500.000	-4.0
Aroclor-1260-3	6.700	6.606	6.806	510.000	500.000	2.0
Aroclor-1260-4	7.170	7.077	7.277	516.710	500.000	3.3
Aroclor-1260-5	7.413	7.319	7.519	510.190	500.000	2.0
Decachlorobiphenyl	8.813	8.723	8.923	52.310	50.000	4.6
Tetrachloro-m-xylene	3.795	3.691	3.891	47.750	50.000	-4.5