

Surrogate Summary

SDG No.: **Q2135**

Client: **ATC Group Services LLC**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO111057.D	PIBLK-PO111057.D	Tetrachloro-m-xylene	1	20	16.4	82		60	140
		Decachlorobiphenyl	1	20	17.5	87		60	140
		Tetrachloro-m-xylene	2	20	15.4	77		60	140
		Decachlorobiphenyl	2	20	17.9	90		60	140
I.BLK-PO111359.D	PIBLK-PO111359.D	Tetrachloro-m-xylene	1	20	21.1	105		60	140
		Decachlorobiphenyl	1	20	20.6	103		60	140
		Tetrachloro-m-xylene	2	20	19.1	96		60	140
		Decachlorobiphenyl	2	20	19.9	100		60	140
PB168212BL	PB168212BL	Tetrachloro-m-xylene	1	20	23.0	115		32	144
		Decachlorobiphenyl	1	20	21.4	107		32	175
		Tetrachloro-m-xylene	2	20	20.5	102		32	144
		Decachlorobiphenyl	2	20	19.9	99		32	175
PB168212BS	PB168212BS	Tetrachloro-m-xylene	1	20	21.2	106		32	144
		Decachlorobiphenyl	1	20	20.9	105		32	175
		Tetrachloro-m-xylene	2	20	19.2	96		32	144
		Decachlorobiphenyl	2	20	19.4	97		32	175
I.BLK-PO111372.D	PIBLK-PO111372.D	Tetrachloro-m-xylene	1	20	22.4	112		60	140
		Decachlorobiphenyl	1	20	20.8	104		60	140
		Tetrachloro-m-xylene	2	20	20.2	101		60	140
		Decachlorobiphenyl	2	20	19.3	96		60	140
I.BLK-PP072163.D	PIBLK-PP072163.D	Tetrachloro-m-xylene	1	20	16.3	82		60	140
		Decachlorobiphenyl	1	20	16.7	84		60	140
		Tetrachloro-m-xylene	2	20	16.6	83		60	140
		Decachlorobiphenyl	2	20	17.0	85		60	140
I.BLK-PP072501.D	PIBLK-PP072501.D	Tetrachloro-m-xylene	1	20	18.9	94		60	140
		Decachlorobiphenyl	1	20	19.0	95		60	140
		Tetrachloro-m-xylene	2	20	20.9	105		60	140
		Decachlorobiphenyl	2	20	21.9	110		60	140
Q2135-01	01A-01B-01C	Tetrachloro-m-xylene	1	20	20.8	104		32	144
		Decachlorobiphenyl	1	20	18.4	92		32	175
		Tetrachloro-m-xylene	2	20	20.7	103		32	144
		Decachlorobiphenyl	2	20	19.2	96		32	175
I.BLK-PP072516.D	PIBLK-PP072516.D	Tetrachloro-m-xylene	1	20	19.0	95		60	140
		Decachlorobiphenyl	1	20	19.7	99		60	140
		Tetrachloro-m-xylene	2	20	20.2	101		60	140
		Decachlorobiphenyl	2	20	22.7	114		60	140