



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2135</u>	Analytical Method:	<u>8082A</u>
Client:	<u>ATC Group Services LLC</u>	Datafile :	<u>PO111361.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB168212BS (Column 1)	AR1016	166.6	143	ug/kg	86				71	120		
	AR1260	166.6	143	ug/kg	86				65	130		
	Total PCBs	0	251	ug/kg					0	0		
PB168212BS (Column 2)	AR1016	166.6	128	ug/kg	77				71	120		
	AR1260	166.6	123	ug/kg	74				65	130		
	Total PCBs	0	286	ug/kg					0	0		