



**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Kinderprep **SDG No.:** Q2145  
**Contract:** KIND02 **Lab Code:** CHEM **Case No.:** Q2145 **SAS No.:** Q2145  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** PLASMA-PURE

---

<b>Sample ID</b>	<b>Analyte</b>	<b>Result ug/L</b>	<b>True Value</b>	<b>% Recovery</b>	<b>Acceptance Window (%R)</b>	<b>M</b>	<b>Analysis Date</b>	<b>Analysis Time</b>	<b>Run Number</b>
ICV01	Copper	106	100	106	90 - 110	P	06/09/2025	17:10	LB136077
	Lead	202	200	101	90 - 110	P	06/09/2025	17:10	LB136077

---



**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Kinderprep **SDG No.:** Q2145  
**Contract:** KIND02 **Lab Code:** CHEM **Case No.:** Q2145 **SAS No.:** Q2145  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** PLASMA-PURE

---

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	5160	5000	103	90 - 110	P	06/09/2025	17:42	LB136077
	Lead	2640	2500	106	90 - 110	P	06/09/2025	17:42	LB136077
CCV02	Copper	4980	5000	100	90 - 110	P	06/09/2025	18:18	LB136077
	Lead	2530	2500	101	90 - 110	P	06/09/2025	18:18	LB136077
CCV03	Copper	5070	5000	101	90 - 110	P	06/09/2025	19:00	LB136077
	Lead	2420	2500	97	90 - 110	P	06/09/2025	19:00	LB136077
CCV04	Copper	5160	5000	103	90 - 110	P	06/09/2025	19:14	LB136077
	Lead	2390	2500	96	90 - 110	P	06/09/2025	19:14	LB136077

---



**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Kinderprep **SDG No.:** Q2145  
**Contract:** KIND02 **Lab Code:** CHEM **Case No.:** Q2145 **SAS No.:** Q2145  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** PLASMA-PURE

---

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	107	100	107	90 - 110	P	06/11/2025	19:40	LB136122
	Lead	206	200	103	90 - 110	P	06/11/2025	19:40	LB136122

---



**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Kinderprep **SDG No.:** Q2145  
**Contract:** KIND02 **Lab Code:** CHEM **Case No.:** Q2145 **SAS No.:** Q2145  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** PLASMA-PURE

---

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	5030	5000	101	90 - 110	P	06/11/2025	20:02	LB136122
	Lead	2460	2500	98	90 - 110	P	06/11/2025	20:02	LB136122
CCV02	Copper	5120	5000	102	90 - 110	P	06/11/2025	20:42	LB136122
	Lead	2490	2500	100	90 - 110	P	06/11/2025	20:42	LB136122
CCV03	Copper	5150	5000	103	90 - 110	P	06/11/2025	21:16	LB136122
	Lead	2470	2500	99	90 - 110	P	06/11/2025	21:16	LB136122
CCV04	Copper	5170	5000	104	90 - 110	P	06/11/2025	21:29	LB136122
	Lead	2460	2500	98	90 - 110	P	06/11/2025	21:29	LB136122

---