



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Kinderprep **SDG No.:** Q2145
Contract: KIND02 **Lab Code:** CHEM **Case No.:** Q2145 **SAS No.:** Q2145
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI	Copper	2.18	2.0	109	70 - 130	P	06/09/2025	17:48	LB136077
	Lead	0.87	1.0	87	70 - 130	P	06/09/2025	17:48	LB136077
CRI	Copper	1.59	2.0	80	70 - 130	P	06/11/2025	20:11	LB136122
	Lead	1.24	1.0	124	70 - 130	P	06/11/2025	20:11	LB136122