



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kinderprep SDG No.: Q2145
Contract: KIND02 Lab Code: CHEM Case No.: Q2145 SAS No.: Q2145

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Copper	2.00	+/-2.00	U*	2.00	P	06/09/2025	17:13	LB136077
	Lead	1.00	+/-1.00	U*	1.00	P	06/09/2025	17:13	LB136077

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Kinderprep</u>		SDG No.: <u>Q2145</u>							
Contract:	<u>KIND02</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2145</u>	SAS No.:	<u>Q2145</u>		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Copper	2.00	+/-2.00	U*	2.00	P	06/09/2025	17:45	LB136077
	Lead	1.00	+/-1.00	U*	1.00	P	06/09/2025	17:45	LB136077
CCB02	Copper	2.00	+/-2.00	U*	2.00	P	06/09/2025	18:24	LB136077
	Lead	1.00	+/-1.00	U*	1.00	P	06/09/2025	18:24	LB136077
CCB03	Copper	1.16	+/-2.00	J*	2.00	P	06/09/2025	19:03	LB136077
	Lead	0.28	+/-1.00	J	1.00	P	06/09/2025	19:03	LB136077
CCB04	Copper	1.18	+/-2.00	J*	2.00	P	06/09/2025	19:17	LB136077
	Lead	1.00	+/-1.00	U*	1.00	P	06/09/2025	19:17	LB136077

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Kinderprep</u>				SDG No.:	<u>Q2145</u>			
Contract:	<u>KIND02</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>Q2145</u>	SAS No.:	<u>Q2145</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Copper	2.00	+/-2.00	U*	2.00	P	06/11/2025	19:43	LB136122
	Lead	1.00	+/-1.00	U*	1.00	P	06/11/2025	19:43	LB136122

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Kinderprep</u>		SDG No.: <u>Q2145</u>							
Contract: <u>KIND02</u>		Lab Code: <u>CHEM</u>		Case No.: <u>Q2145</u>		SAS No.: <u>Q2145</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Copper	2.00	+/-2.00	U*	2.00	P	06/11/2025	20:05	LB136122
	Lead	0.69	+/-1.00	J*	1.00	P	06/11/2025	20:05	LB136122
CCB02	Copper	2.00	+/-2.00	U*	2.00	P	06/11/2025	20:45	LB136122
	Lead	0.29	+/-1.00	J	1.00	P	06/11/2025	20:45	LB136122
CCB03	Copper	2.00	+/-2.00	U*	2.00	P	06/11/2025	21:19	LB136122
	Lead	0.56	+/-1.00	J*	1.00	P	06/11/2025	21:19	LB136122
CCB04	Copper	2.00	+/-2.00	U*	2.00	P	06/11/2025	21:34	LB136122
	Lead	1.00	+/-1.00	U*	1.00	P	06/11/2025	21:34	LB136122

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Kinderprep</u>	SDG No.:	<u>Q2145</u>						
Contract:	<u>KIND02</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2145</u>	SAS No.:	<u>Q2145</u>		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number