

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Kinderprep **SDG No.:** Q2145
Contract: KIND02 **Lab Code:** CHEM **Case No.:** Q2145 **SAS No.:** Q2145
ICS Source: EPA **Instrument ID:** P8

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	9.57	8.0	120	4	12	06/09/2025	17:20	LB136077
	Lead	4.27	4.0	107	2	6	06/09/2025	17:20	LB136077
ICSAB01	Copper	28.9	25.0	116	21	29	06/09/2025	17:37	LB136077
	Lead	22.5	25.0	90	21.3	28.8	06/09/2025	17:37	LB136077
ICSA01	Copper	8.92	8.0	112	4	12	06/11/2025	19:49	LB136122
	Lead	5.41	4.0	135	2	6	06/11/2025	19:49	LB136122
ICSAB01	Copper	28.0	25.0	112	21	29	06/11/2025	19:59	LB136122
	Lead	22.7	25.0	91	21.3	28.8	06/11/2025	19:59	LB136122