

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Kinderprep</u>	level: <u>low</u>	sdg no.: <u>Q2145</u>
contract: <u>KIND02</u>	lab code: <u>CHEM</u>	case no.: <u>Q2145</u> sas no.: <u>Q2145</u>
matrix: <u>Water</u>	sample id: <u>Q2145-14</u>	client id: <u>CLASSROOM-SINK-16MS</u>
Percent Solids for Sample: NA	Spiked ID: Q2145-14MS	Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	70 - 130	5030		31.4		5000	100		P
Lead	ug/L	70 - 130	2190		1.00	U	2500	88		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Kinderprep</u>	level: <u>low</u>	sdg no.: <u>Q2145</u>
contract: <u>KIND02</u>	lab code: <u>CHEM</u>	case no.: <u>Q2145</u> sas no.: <u>Q2145</u>
matrix: <u>Water</u>	sample id: <u>Q2145-14</u>	client id: <u>CLASSROOM-SINK-16MSD</u>
Percent Solids for Sample: NA	Spiked ID: Q2145-14MSD	Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	70 - 130	5200		31.4		5000	103		P
Lead	ug/L	70 - 130	2310		1.00	U	2500	92		P

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Kinderprep **level:** low **sdg no.:** Q2145
contract: KIND02 **lab code:** CHEM **case no.:** Q2145 **sas no.:** Q2145
matrix: Water **sample id:** Q2145-31 **client id:** CLASSROOM-SINK-16 -BMS
Percent Solids for Sample: NA **Spiked ID:** Q2145-31MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	70 - 130	5260		126		5000	103		P
Lead	ug/L	70 - 130	2410		0.40	J	2500	96		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Kinderprep **level:** low **sdg no.:** Q2145
contract: KIND02 **lab code:** CHEM **case no.:** Q2145 **sas no.:** Q2145
matrix: Water **sample id:** Q2145-31 **client id:** CLASSROOM-SINK-16 -BMSD
Percent Solids for Sample: NA **Spiked ID:** Q2145-31MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	70 - 130	5260		126		5000	103		P
Lead	ug/L	70 - 130	2390		0.40	J	2500	95		P